

STATE FORENSIC SCIENCE LABORATORY

5 CIVIL LINES, SAGAR (M.P.) 470 001

{Ph No.- 07582-267707 Fax No.- 07582-267791 e-mail : dir_fsl@mppolice.gov.in}

क्रमांक / एफ.एस.एल. / स्टोर्स / सीनि / डीएनए / एसबी18-19 / 2019 दिनांक 05 / 02 / 2019

// सीमित निविदा आमंत्रण की सूचना //

राज्य बजट वर्ष 2018-19 के अंतर्गत तालिकानुसार किट्स/रसायनों का क्रय किया जाना है, जिनके तकनीकी स्पेसिफिकेशन संलग्न है। निम्न शर्तों के अधीन दिनांक 15/02/2019 तक सीमित निविदा प्रस्तुत की जा सकती है :-

1. प्रत्येक आयटम हेतु अलग-अलग सीमित निविदा प्रस्तुत करें।
2. निविदा में सामग्री की बेसिक कीमत व जीएसटी का पृथक-पृथक उल्लेख करें।
3. राज्य न्यायालयिक विज्ञान प्रयोगशाला डीएसएआईआर में **Notification no. 51/1996- Customs, Dated 23.07.1996** के अंतर्गत रजिस्टर्ड है। अतः इसके अंतर्गत जीएसटी 5% ही देय होगा।
4. सीमित निविदा हेतु दिये गये लिफाफे पर संदर्भ क्रमांक/एफ.एस.एल./स्टोर्स/सीनि/डीएनए/एसबी18-19/2019 दिनांक 05/02/2019 व सामग्री का नाम लेख करें।
5. वेलिडिटी कम से कम 180 दिवस देना अनिवार्य है।
6. सामग्री का मेक/माडल, पम्फलेट्स, ब्रोसर्स इत्यादि आवश्यक रूप से संलग्न करें।
7. उपकरण की गारंटी 01 वर्ष अनिवार्य है।
8. वर्तमान वित्तीय वर्ष का जीएसटी जमा करने का सर्टिफिकेट, यूजर लिस्ट तथा फर्म रजिस्ट्रेशन एवं फर्म की वित्तीय स्थिति की जानकारी की सत्यापित फोटोप्रति संलग्न करें।
9. ईएमडी - निदेशक एफएसएल के नाम से बैंक ड्राफ्ट, बैंक गारंटी/एफडीआर (वैधता अवधि 06 माह) संलग्न करें।
10. सामग्री की सप्लाई एफओआर बेसिस पर की जायेगी।
11. देयकों का भुगतान सामग्री के प्रदाय व समिति द्वारा वेरिफिकेशन/इन्स्टालेशन के पश्चात् किया जावेगा।

List of Items

S.No.	DNA Kits/Consumables	QTY	Earnest Money Rs.
1	Genotyping Kit for Autosomal STR Markers	200 Reactions	4000.00
2	Genotyping Kit for Y chromosome Markers	100 Reactions	1000.00

TECHNICAL SPECIFICATION

1. Genotyping Kit for Autosomal STR Markers

- Kit should have the minimum 20 CODIS Autosomal STR markers and amelogenin with published validation on the kit in PUBMED JOURNAL.
- Kit should work on 6 dye chemistry and have built-in features such as resistance to PCR inhibitors.
- Kit should have one or more Y STR/INDEL marker
- Kit should work on ABI 3500XL Genetic analyzer.
- Kit sensitivity- equal to or less than 500 picogram
- Kit should be supplied with appropriate quantity of Control DNA, Size standard, Allelic ladder and PCR grade water (cost of Control DNA, Size standard, PCR grade water and Allelic ladder should be included in the kit)
- Kit should be supplied in small aliquots of 100 or 200 reactions
- Suitable Matrix standards one pack should be supplied with this supply order

2. Genotyping Kit for Y chromosome Markers

- Kit should have the minimum 20 Y-STR markers with published validation on the kit in PUBMED JOURNAL.
- Kit should work on 5/6 dye chemistry and have built-in features such as resistance to PCR inhibitors.
- Kit should work on ABI 3500XL Genetic analyzer.
- Kit sensitivity- equal to or less than 200 picograms
- Kit should be supplied with appropriate quantity of Control DNA, size standard, allelic ladder and PCR grade water (cost of Control DNA, size standard, PCR grade water and allelic ladder should be included in the kit)
- Suitable Matrix standards should be supplied with this supply order.
- Y-STR kit for specific and sensitive detection of male DNA in the presence of high amounts of female DNA
- Kit should be supplied in small aliquots of 100 or 200 reactions

निदेशक
राज्य न्यायालयिक विज्ञान प्रयोगशाला,
सागर म.प्र.